


PULSE GENERATOR
PERFORMANCE CHECK

Model: APD-3A-C-P

S.N.: 4932

Date: JUNE 16 '89



- a) Output signal amplitude: +25 to +75 AMPS
- b) Pulse width: 2.5 nSEC (HART POWER)
- c) Rise time: NA
- d) Fall time: NA
- e) PRF: 10 TO 20 KHz
- f) Jitter, stability: OK
- g) Prime power: 120 / 240 V
50 - 60 Hz

$R_L = 5 \Omega$
60 db ATTEN
 $\therefore = 20 \text{ AMPS/DIV}$
PRF = 20 KHz

[Signature]