


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVMP-2-C-PN-M-OT*
S.N.: *4968*
Date: *JUL 7 14 '89*



40 db ATTEN
10 VOLTS/DIV
PRF = 1.0 MHz

- a) Output signal amplitude:
0 TO ± 10 VOLTS
- b) Pulse width:
5 TO 100 NSER
- c) Rise time:
≤ 100 PSEK
- d) Fall time:
≤ 135 PSEK
- e) PRF:
0 TO 1.0 MHz
- f) Jitter, stability:
OK
- g) Prime power:
120/240 V
50-60 Hz