


PULSE GENERATOR  
PERFORMANCE CHECK

Model: AV6-1-P  
S.N.: 5046  
Date: July 19 89



60 dB ATTEN  
100 VOLTS/DIV  
PRF = 20 KHz

- a) Output signal amplitude:  
0 TO +250 VOLTS
- b) Pulse width:  
≤ 4 NSEC
- c) Rise time:  
NA
- d) Fall time:  
NA
- e) PRF:  
0 TO 50 KHz
- f) Jitter, stability:  
OK
- g) Prime power:  
+15 VOLT  
250 mA (MAX)

*[Signature]*