

PULSE GENERATOR  
PERFORMANCE CHECK

Model: AP-M-4U2-C-P-05

S.N.: 5070

Date: AUG 8 1989



40 dB ATTEN

1000003 1110

PRF  $\approx$  100 KHz

- a) Output signal amplitude:  
0 TO +20 VOLTS
- b) Pulse width:  
0.3 TO 2.0 NS
- c) Rise time:  
 $\leq$  100 PSEC
- d) Fall time:  
 $\leq$  200 PSEC
- e) PRF:  
0 TO 1.0 MHz
- f) Jitter, stability:  
OK
- g) Prime power:  
120/240 V  
50-60 Hz

AL