


PULSE GENERATOR
PERFORMANCE CHECK

Model: AVO-2 PN

S.N.: 5094

Date: JULY 20 1989



≈ 6 AMPS/DIV

PRF = 20 KHz

- a) Output signal amplitude:
0 TO ± 10 AMPS
- b) Pulse width:
2 TO 15 NS
- c) Rise time:
≤ 1 NS
- d) Fall time:
≤ 2 NS
- e) PRF:
0 TO 20 KHz
- f) Jitter, stability:
OK
- g) Prime power:

+15 ± 1.0 VDC
300 mA (MAX)

[Signature]