


PULSE GENERATOR  
PERFORMANCE CHECK

Model: *PVI-100P-P*

S.N.: *5104*

Date: *JULY 31 1949*



*56 db ATTEN*  
*32 VOLT/DIV*  
*PRF = 50 KHZ*

- a) Output signal amplitude:  
*7 + 40 VOLTS*
- b) Pulse width:  
*2 TO 100 NSER*
- c) Rise time:  
*≤ 350 PSEC*
- d) Fall time:  
*≤ 2 NSER*
- e) PRF:  
*MIN PW: 0 TO 10 KHZ*  
*MAX PW: 0 TO 20 KHZ*
- f) Jitter, stability:  
*SEE FIG 1+2*
- g) Prime power:  
*OK*  
*+15 ± 1.0 VOLT*  
*150 mA (MAX)*

*[Signature]*