


PULSE GENERATOR

PERFORMANCE CHECK

Model: *AVL-2A-C-W-N*

S.N.: *5115*

Date: *OCT 18 1989*



60 dB ATTEN

100 VOLTS/DIV

PRF = 4.0 KHz

a) Output signal amplitude:

0 TO - 170 VOLTS

b) Pulse width:

5 TO 400 NS *

c) Rise time:

≤ 2

d) Fall time:

≤ 3 NS

e) PRF:

0 TO 5 KHz

f) Jitter, stability:

OK

g) Prime power:

120/240 V

50-60 Hz

* *NEEDS ADDITIONAL EXTERNAL CABLE FOR PW RANGE 100 TO 400 NS. SEE NOTE 5.*