


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVO-SB-C-P-EP-EW-M*

S.N.: *5116*

Date: *JUNE 28 1989*



- a) Output signal amplitude:
0 TO +40.0mV
- b) Pulse width:
25 TO 500 nsec
- c) Rise time:
≤ 10 nsec
- d) Fall time:
≤ 10 nsec
- e) PRF:
0 TO 1 kHz
- f) Jitter, stability:
OK
- g) Prime power:

10 VOLTS/DIV
R_i = 0.7Ω
PRF = 1.0 kHz

120/240V
50-60Hz