


PULSE GENERATOR  
PERFORMANCE CHECK

Model: *APR-1-PW-PS-P*

S.N.: *5139*

Date: *JUN 7 31 1989*



*R<sub>L</sub> = 50  $\Omega$*   
*f<sub>REP</sub> = 1.0 KHz*

- a) Output signal amplitude:  
*0 TO +200 VOLTS*
- b) Pulse width:  
*0.1 TO 5.0  $\mu$ SEC*
- c) Rise time:  
*500 NS*
- d) Fall time:  
*510 NS*
- e) PRF:  
*0 TO 1 KHz*
- f) Jitter, stability:  
*OK*
- g) Prime power:  
*120/240 V*  
*50-60 Hz*

*[Signature]*