


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVMM-2-P*

S.N.: *5142*

Date: *JUN 9 31 1989*



40 db ATTEN
10 VOLT/DIV
PRF = 12 MHz

- a) Output signal amplitude:
0 TO +15V
- b) Pulse width:
0.2 TO 2.0 nSEC
- c) Rise time:
≤ 100 pSEC
- d) Fall time:
≤ 135 pSEC
- e) PRF:
0 TO 25 MHz
- f) Jitter, stability:
OK
- g) Prime power:
+24 ± 1.0 VOLT
500 mA (MAX)
[Signature]