


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVD-AU-1-C-P*

S.N.: *5168*

Date: *Aug 25 1969*



40 dB ATTEN
-10 VOLTS/DIV
PRF = 100 KHz

- a) Output signal amplitude:
0 TO +10 VOLTS
- b) Pulse width:
0.2 TO 4.0 nSEC
- c) Rise time:
≤ 100 pSEC
- d) Fall time:
≤ 135 pSEC
- e) PRF:
0 TO 1.0 MHz
- f) Jitter, stability:
OK
- g) Prime power:
120/240 V
50-60 Hz.

[Signature]