


PULSE GENERATOR

PERFORMANCE CHECK

Model: AVH-5-MV2-P-05A

S.N.: 5184

Date: OCT 31 1989



~ 40 dB ATTEN  
10 VOLTS/DIV  
PRF = 500 KHz

a) Output signal amplitude:

7 + 30 VOLTS

b) Pulse width:

≤ 350 ps

c) Rise time:

NA

d) Fall time:

NA

e) PRF:

0 TO 1 MHz

f) Jitter, stability:

OK

g) Prime power:

+15 ± 1.0 VOLTS

≤ 200 mA