


PULSE GENERATOR
PERFORMANCE CHECK

Model: *APG-B2 P-W-CA-CA*

S.N.: *5185*

Date: *Aug 16 89*



$R_L = 50 \Omega$

PRF = 10 KHz

a) Output signal amplitude:

0 TO +50 VOLTS

b) Pulse width:

0.1 TO 5.0 μ SEC

c) Rise time:

≤ 5 NSEC

d) Fall time:

≤ 5 NSEC

e) PRF:

0 TO 20 KHz

f) Jitter, stability:

OK

g) Prime power:

+24 \pm 1.0 VOLTS

500 mA MAX