



PULSE GENERATOR
PERFORMANCE CHECK

Model: *PMMP-2-N*
S.N.: *5219*
Date: *NOV 8 89*



(A) 40 dB ATTEN
10 VOUT/DIV
PRF = 500 KHz



(B) (RISE TIME)
40 dB ATTEN
10 VOUT/DIV
PRF = 500 KHz

- a) Output signal amplitude:
0 TO -10 V
- b) Pulse width:
5 TO 100 nSEC
- c) Rise time:
≤ 100 pSEC
- d) Fall time:
≤ 100 pSEC
- e) PRF:
0 TO 1.0 MHz
- f) Jitter, stability:
OK
- g) Prime power:
+24 ± 1.0 VDC
250 mA (MAX)