


PULSE GENERATOR

PERFORMANCE CHECK

Model: AVH-MPT-7

S.N.: 5227

Date: NOV 15' 1989



a) Output signal amplitude:

$\geq +30$ VOLTS.

b) Pulse width:

≤ 1.0 NSEC

c) Rise time:

NA

d) Fall time:

NA

e) PRF:

0 TO 1.0 MHz

f) Jitter, stability:

OK

g) Prime power:

$+15 \pm 1.0$ VOLTS.
150 mA (MAX)

40 dB ATTEN

10 VOLTS/DIV

PRF = 700 kHz