


PULSE GENERATOR  
PERFORMANCE CHECK

Model: AVL-2-C-PN-05-M  
S.N.: 5244  
Date: NOV 24 1989



70 66 ATTEN  
≈ 320 VOLTS/DIV  
PRF = 1.0/KHZ

- a) Output signal amplitude:  
0 TO ±350 VOLTS
- b) Pulse width:  
5 TO 100 NS
- c) Rise time:  
6.2 NS
- d) Fall time:  
5.3 NS
- e) PRF:  
0 TO 5 KHZ
- f) Jitter, stability:  
OK
- g) Prime power:  
120/240 VOLTS  
50-60 HZ