


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVM-1-e-PN-D*

S.N.: *5259*

Date: *DEC 1 89*



40 db ATTEN
5 V0 250 / DIV
PRF \approx 12 MHz.

- a) Output signal amplitude:
0 TO 55 VOLTS
- b) Pulse width:
0.2 TO 6 NSEC
- c) Rise time:
 \leq 100 PSEC
- d) Fall time:
 \leq 150 PSEC
- e) PRF:
0 TO 25 MHz.
- f) Jitter, stability:
OK
- g) Prime power:
120 / 240 V
50 - 60 Hz