


PULSE GENERATOR
PERFORMANCE CHECK

Model: AVMR-1-P-EA-EW-EO-TRF-ECL-MI-JM

S.N.: 5263

Date: NOV 27/89



40 db ATTEN
∴ 5 VOLTS/DIV
7.0 MHz PRF

- a) Output signal amplitude:
0 TO +5 VOLTS
- b) Pulse width:
5 TO 50 NSEC
- c) Rise time:
 ≤ 100 PSEC OR ≤ 1 NSEC
- d) Fall time:
 ≤ 135 PSEC OR ≤ 1 NSEC
- e) PRF:
0 TO 10 MHz
- f) Jitter, stability:
OK
- g) Prime power:
+24 \pm 1.0 VOLT
300 mA (MAX)