



PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVK-4A-PW-CPN-CAEW-CNR1*
 S.N.: *5282*
 Date: *JUNE 15 90*



- a) Output signal amplitude:
0 TO ± 400 V TO
- b) Pulse width: *R_L > 10K*
0.2 TO 2.0 μSEC
- c) Rise time:
≤ 10 nSEC
- d) Fall time:
≤ 10 nSEC
- e) PRF: *0 TO 20 kHz*
- f) Jitter, stability:
OK
- g) Prime power:
170/240 V
50-60 Hz

P_{out} R_L = 10K
PRF = 20 kHz



- h) Prop Delay:
≤ 50 nSEC
(x 33 nSEC)

Wont R_L = 10K
PRF = 20 kHz