


PULSE GENERATOR

PERFORMANCE CHECK

Model: *AWL-2-PS-N-CA-EW*

S.N.: *5285*

Date: *NOV 21 1989*



*70 dB ATTEN*

*≈ 320 VOLTS/DIV*

*PRF = 2 KHz*

a) Output signal amplitude:

*0 TO - 350 VOLTS*

b) Pulse width:

*5 TO 100 NSER*

c) Rise time:

*≤ 2 NSER*

d) Fall time:

*≤ 3 NSER*

e) PRF:

*0 TO 5 KHz*

f) Jitter, stability:

*OK*

g) Prime power:

*120/240V*

*50-60 Hz*