



PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVM-1-C-N-PN-OT*
S.N.: *5301*
Date: *JAN 10 90*



*(A) 40dB ATTEN
5 VOLTS/DIV
PRF ≈ 12 MHz*



*(B) 40dB ATTEN
5 VOLTS/DIV
PRF ≈ 25 MHz*

- a) Output signal amplitude:
0 TO 7.5 VOLTS
- b) Pulse width:
0.2 TO 6.0 NSER
- c) Rise time:
≤ 100 PSEK
- d) Fall time:
≤ 135 PSEK
- e) PRF:
0 TO 25 MHz
- f) Jitter, stability:
OK
- g) Prime power:
*120/240V
50-60Hz*