


PULSE GENERATOR
PERFORMANCE CHECK

Model: PWL-2A-C-W-PN-05

S.N.: 5327

Date: JAN 16 '90



60 db ATTEN
100 VOLTS/DIV
PRF = 5 KHZ

- a) Output signal amplitude:
0 TO ± 170 VOLTS
- b) Pulse width:
5 TO 400 NS^{*}
- c) Rise time:
≤ 2 NS
- d) Fall time:
≤ 3 NS
- e) PRF:
0 TO 5 KHZ
- f) Jitter, stability:
OK
- g) Prime power:
120/240 V
50-60 Hz

* NEEDS ADDITIONAL
EXTERNAL DELAY
LINE FOR 100
TO 400 NS