


PULSE GENERATOR  
PERFORMANCE CHECK

Model: AVH-HV.3-P

S.N.: 5329

Date: JAN 4 90



50 db ATTEN

32 VOLTS/DIV

PRF = 200 kHz

- a) Output signal amplitude:  
0 TO + 60 VOLTS
- b) Pulse width:  
E 1 NSEC
- c) Rise time:  
NA
- d) Fall time:  
NA
- e) PRF:  
0 TO 1 MHz
- f) Jitter, stability:  
OK
- g) Prime power:  
+ 15 ± 1.0 VOLT  
150 mA (MAX)