


PULSE GENERATOR  
PERFORMANCE CHECK

Model: AVH-PM2-P  
S.N.: 5330  
Date: JAN 4 1990



50 db ATTEN  
32 VOLTS/DIV  
Z RF ≈ 700 KHz.

- a) Output signal amplitude:  
≥ +60 VOLTS
- b) Pulse width:  
≤ 2 NSEC
- c) Rise time:  
NA
- d) Fall time:  
NA
- e) PRF:  
0 TO 1.0 MHz
- f) Jitter, stability:  
OK
- g) Prime power:  
+135 ± 5 VOLTS  
5 mA (MAX)

*[Signature]*