


PULSE GENERATOR  
PERFORMANCE CHECK

Model: AVO-2W-P  
S.N.: 5333  
Date: JAN 29 1990



~ 7 AMPS / DIV  
PRF = 20 KHZ

- a) Output signal amplitude:  
0 TO +10 AMPS
- b) Pulse width:  
2 TO 50 NSER
- c) Rise time:  
≤ 1 NSER
- d) Fall time:  
≤ 2 NSER
- e) PRF:  
0 TO 20 KHZ
- f) Jitter, stability:  
OK
- g) Prime power:  
+15 ± 1.0 VOLTS  
300 mA (MAX)