


PULSE GENERATOR

PERFORMANCE CHECK

Model: *APM-1-C-PN*

S.N.: *5352*

Date: *JAN 19 90*



40 dB ATTEN

5 VOLT/DIV

PRF ≈ 12 MHz

- a) Output signal amplitude:
0 TO ± 5 VOLTS
- b) Pulse width:
0.2 TO 6.0 NS
- c) Rise time:
≤ 100 PSEC
- d) Fall time:
≤ 135 PSEC
- e) PRF:
0 TO 25 MHz
- f) Jitter, stability:
OK
- g) Prime power:
120/240 V
50-60 Hz.

[Signature]