


PULSE GENERATOR  
PERFORMANCE CHECK

Model: AVH-5-1-C-N-05

S.N.: 5353

Date: Feb 19 90



40 dB ATTEN

.10 VOLT/DIV

PRF = 100 KHz

a) Output signal amplitude:

0 TO -10V

b) Pulse width:

≤ 130 pSEC

c) Rise time:

NA

d) Fall time:

NA

e) PRF:

0 TO 1000 Hz

f) Jitter, stability:

OK

g) Prime power:

120 / 240 V

50-60 Hz