


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVL-AV-1-P*

S.N.: *5364*

Date: *FEB 22 90*



60 db ATTEN

100 VOLTS/DIV

PRF = 5 KHz

- a) Output signal amplitude:
0 TO 100 VOLTS
- b) Pulse width:
5 TO 100 NSEC
- c) Rise time:
≤ 1 NSEC
- d) Fall time:
≤ 2 NSEC
- e) PRF:
0 TO 5 KHz
- f) Jitter, stability:
OK
- g) Prime power:
+15 ± 1.0 VOLTS
500 mA (MAX)