


PULSE GENERATOR

PERFORMANCE CHECK

Model: *AVL-AU-1-P*

S.N.: *5365*

Date: *FEB 22 90*



*60 db ATTEN*

*100 VOLTS/DIV*

*PRF = 5 KHZ*

a) Output signal amplitude:

*0 TO +100 VOLTS*

b) Pulse width:

*5 TO 100 NSEC*

c) Rise time:

*≤ 1 NSEC*

d) Fall time:

*≤ 2 NSEC*

e) PRF:

*0 TO 5 KHZ*

f) Jitter, stability:

*OK*

g) Prime power:

*+15 ± 1.0 VOLTS*

*500 mA (MAX)*

*[Signature]*