


PULSE GENERATOR
PERFORMANCE CHECK

Model: AVR-B1-P
S.N.: 5368
Date: JAN 15 90



40 db ATTEN
10 VOLTS/DIV
PRF ≈ 50 KHz

- a) Output signal amplitude:
0 TO +20 VOLTS
- b) Pulse width:
20 TO 200 nSEC
- c) Rise time:
≤ 2 nSEC
- d) Fall time:
≤ 2 nSEC
- e) PRF:
0 TO 100 KHz
- f) Jitter, stability:
OK
- g) Prime power:
+24 ± 1.0 VOLTS
200 mA (max)