


PULSE GENERATOR  
PERFORMANCE CHECK

Model: *AP6-1-P*  
S.N.: *5369*  
Date: *MAR 8 '90*



*60 db ATTEN*  
*100 VOLTS/DIV*  
*PRF = 25 KHZ*

- a) Output signal amplitude:  
*0 TO +250 V*
- b) Pulse width:  
*≤ 4 nSEC*
- c) Rise time:  
*NA*
- d) Fall time:  
*NA*
- e) PRF:  
*0 TO 50 KHZ*
- f) Jitter, stability:  
*OK*
- g) Prime power:  
*+15 ± 1.0 VOLT*  
*≤ 250 mA.*