


PULSE GENERATOR
PERFORMANCE CHECK

Model: ANM-1-C-BP1-M-P

S.N.: 5378

Date: DMAR 9 90



40 db ATTEN

5 VOLTS/DIV

PRF \approx 12 MHz

- a) Output signal amplitude:
0 TO +5 VOLTS
- b) Pulse width:
0.3 TO 10 NSEC
- c) Rise time:
 \leq 100 PSEC
- d) Fall time:
 \leq 150 PSEC
- e) PRF: 0 TO 25 MHz
- f) Jitter, stability:
OK
- g) Prime power:
120/240 V
50-60 Hz

[Handwritten signature]