



PULSE GENERATOR
PERFORMANCE CHECK

Model: *MMMP-3-CP*
S.N.: *5390*
Date: *MAR 7 90*



Ⓐ *40 db ATTEN*
10 VOLTS/DIV
PRF = 1.0 MHz



Ⓑ *RISE TIME*

- a) Output signal amplitude:
0 TO +20 VOLTS
- b) Pulse width:
8 TO 100 NS
- c) Rise time:
≤ 200 PSEZ
- d) Fall time:
≤ 300 PSEZ
- e) PRF:
0 TO 1.0 MHz
- f) Jitter, stability:
OK
- g) Prime power:
120/200 V
50-60 W