


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVM-2-C-P-PN-D*

S.N.: *5392*

Date: *MAR 22 '90*



- a) Output signal amplitude:
0 TO ±15 VOLTS
- b) Pulse width:
0.2 TO 2.0 NS
- c) Rise time:
≤ 100 PSEC
- d) Fall time:
≤ 135 PSEC
- e) PRF:
0 TO 25 MHz
- f) Jitter, stability:
OK
- g) Prime power:
120/240V
5.0-6.0 Hz

40 db atten
10 VOLTS/DIV
PRF ≈ 12 MHz