


PULSE GENERATOR  
PERFORMANCE CHECK

Model: *AD-7C-C-PN-CA-EW*


S.N.: *5398*

Date: *APRIL 18 1990*



- a) Output signal amplitude:  
*0 TO ±30 AMPS*
- b) Pulse width: *(0 TO 30 μS)*  
*0.2 TO 50 μS*
- c) Rise time:  
*≤ 50 NS*
- d) Fall time:  
*≤ 50 NS*
- e) PRF:  
*MAX PW: 0 TO 50 Hz*  
*PW ≤ 2.5 μS: 0 TO*
- f) Jitter, stability:  
*1 kHz*

*P<sub>OUT</sub> R<sub>L</sub> = 1.0 Ω*  
*10 VOLTS/DIV*  
*0.5 μS/DIV*



g) Prime power: *OK*  
*120/240 V*  
*50-60 Hz*

*P<sub>OUT</sub> R<sub>L</sub> = 1.0 Ω*  
*10 VOLTS/DIV*  
*10 μS/DIV*

*[Signature]*