


PULSE GENERATOR  
PERFORMANCE CHECK

Model: *AVK-1-PW-PS-N*

S.N.: *5405*

Date: *MAR 13 1990*



*50 VOLTS/DIV*  
*0.5 USEC/DIV*  
*R<sub>L</sub> = 50 Ω*  
*PRF = 10 KHz*

- a) Output signal amplitude:  
*0 TO -200 VOLTS*
- b) Pulse width: *TO 50 n*
- c) Rise time:  
*0.1 TO 50 USEC*
- d) Fall time:  
*≤ 10 USEC*
- e) PRF:  
*0 TO 1 KHz.*
- f) Jitter, stability:  
*OK*
- g) Prime power:  
*120 / 240 VOLTS*  
*50 - 60 Hz*