


PULSE GENERATOR  
PERFORMANCE CHECK

Model: *AVN-1-P*  
S.N.: *5408*  
Date: *APRIL 20 1990*



*40 db ATTEN*  
*5 VOLT/DIV*  
*PRF ≈ 200 MHz*

- a) Output signal amplitude:  
*0 TO + 5 VOLTS*
  - b) Pulse width:  
*0.2 TO 1.0 NS*
  - c) Rise time:  
*≤ 100 PSEZ*
  - d) Fall time:  
*≤ 100 PSEZ*
  - e) PRF:  
*50 - 250 MHz*
  - f) Jitter, stability:  
*OK*
  - g) Prime power:  
*+ 24 ± 1.0 VOLT*  
*400 mA (max)*
- [Signature]*