


PULSE GENERATOR

PERFORMANCE CHECK

Model: *MP-M-15-P-PW-ED-C*


S.N.: *5410*

Date: *NOV 15 90*




- a) Output signal amplitude:  
*0 TO ±5 VOLTS*
- b) Pulse width:  
*0.2 TO 4.0 NSEC*

*Ⓐ Post PRF = 100 KHz*




- c) Rise time:  
*≤ 60 PSEC*
- d) Fall time:  
*≤ 135 PSEC*
- e) PRF:  
*0 TO 1.0 MHz*

*Ⓑ Post (RISE TIME)*



- f) Jitter, stability:  
*OK*
- g) Prime power:  
*120/240 V*  
*50-60 Hz*

*Ⓒ Next*



*A, B, C, D*  
*40 dB ATTEN*  
*5 VOLTS/DIV*  
*PRF = 100 KHz*

*Ⓓ Next (RISE TIME)*