


PULSE GENERATOR  
PERFORMANCE CHECK

Model: *AVR-3-C-N-WSU1*

S.N.: *5419*

Date: *APRIL 30 1990*



- a) Output signal amplitude: *0 TO -200 VOLTS*
- b) Pulse width: *TO  $R_L \geq 10K$ .  
0.5  $\mu$ SEC TO 0.5 MSEC*
- c) Rise time:  *$\leq 10$  NSEC*
- d) Fall time:  *$\leq 10$  NSEC*
- e) PRF: *1 KHZ TO 1 MHz*
- f) Jitter, stability: *OK*
- g) Prime power: *120/240 V  
50-60 Hz.*
- h) DUTY CYCLE: *FIXED AT  
50 %*

*50 VOLTS/DIV  
1.0  $\mu$ SEC/DIV  
 $R_L = 100K$ .*