


PULSE GENERATOR  
PERFORMANCE CHECK

Model: PVH-M1-P  
S.N.: 5432  
Date: APRIL 3 90



- a) Output signal amplitude:  $\approx +30$  VOLTS
- b) Pulse width:  $\leq 1.0$  NSEC.
- c) Rise time: NA
- d) Fall time: NA
- e) PRF: 0 TO 1.0 MHz.
- f) Jitter, stability: OK
- g) Prime power:  $+135V \pm 5.0 V$   
5 mA MAX

50 db ATTEN  
32 VOLTS / DIV  
PRF = 700 kHz