


PULSE GENERATOR

PERFORMANCE CHECK

Model: *AMP-AD-15-C-N*

S.N.: *5445*

Date: *APRIL 24 1990*



40 db ATTEN
5 VOLTS/DIV
PRF = 10 MHz

- a) Output signal amplitude:
0 TO -5 VOLTS
- b) Pulse width:
0.2 TO 40 NSEC
- c) Rise time:
≤ 60 PSEC
- d) Fall time:
≤ 135 PSEC
- e) PRF:
0 TO 1.0 MHz
- f) Jitter, stability:
OK
- g) Prime power:
120/240V
50-60Hz