


PULSE GENERATOR
PERFORMANCE CHECK

Model: AVR-A-1-PW-C-PN

S.N.: 5502

Date: MAY 23 1990



a) Output signal amplitude:
0 TO ± 200 VOLTS

b) Pulse width:
50 NSEC TO 1.0 USEC

c) Rise time:
 ≤ 10 NSEC


d) Fall time:
 ≤ 10 NSEC

e) PRF: 0 TO 100 KHZ
0.5% MAX DUTY

f) Jitter, stability: Cycle
OK

g) Prime power:
120/240V
50-60KHZ

P_{out} $R_L = 50 \Omega$
200 NSEC/DIV
5 KHZ PKF



P_{out} $R_L = 50 \Omega$
100 NSEC/DIV
100 KHZ