

PULSE GENERATOR  
PERFORMANCE CHECK

Model: *AVH-141-P-55A*

S.N.: *5505*

Date: *MAY 31 1990*



*50 db ATTEN*

*~ 32 VOLTS/PIN*

*PRF = 70 KHz*

a) Output signal amplitude:  
*> 100 VOLTS TO*

b) Pulse width:  
*50ns*  
*≤ 10 NSEC.*

c) Rise time:  
*NA*

d) Fall time:  
*NA*

e) PRF:  
*0 TO 100 KHz*

f) Jitter, stability:  
*OK*

g) Prime power:  
*+15 VOLT*  
*150 mA (MAX)*