


PULSE GENERATOR
PERFORMANCE CHECK

Model: *ACR-1-PW-C-2*
S.N.: *5506*
Date: *MAY 22 1990*



- a) Output signal amplitude:
0 TO +200 VOLTS
- b) Pulse width:
0.1 TO 5.0 USEC
- c) Rise time:
≤ 10 NSEC
- d) Fall time:
≤ 10 NSEC
- e) PRF:
0 TO 1 K/HZ
- f) Jitter, stability:
OK
- g) Prime power:
120/240 V
50 to 60 HZ

R_i = 50 Ω
50 VOLTS/DIV
200 NSEC/DIV
PRF = 1.0 K/HZ