

PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVP-AV-1-C-LIB*

S.N.: *5512*

Date: *MAY 29 1990*



a) Output signal amplitude:

0 TO +5.0 VOLTS

b) Pulse width: A) *0.2 TO 4.0 NS*

B) *0.2 TO 4.0 NS*

c) Rise time: C) *0.3 TO 4.0 NS*

≤ 200 PSEZ

d) Fall time:

≤ 200 PSEZ.

e) PRF:

0 TO 1.0 MHz

f) Jitter, stability:

OK

g) Prime power:

120/240 V

50-60 Hz.

40 db ATTEN

5 VOLTS/DIV

1.0 MHz