


PULSE GENERATOR
PERFORMANCE CHECK

Model: *A20-3A-P*
 S.N.: *5519*
 Date: *JUNE 8 1990*



- a) Output signal amplitude: *+25 TO +75 AMPS*
- b) Pulse width: *2.0 ± 0.5 NS*
- c) Rise time: *NA*
- d) Fall time: *NA*
- e) PRF: *0 TO 20 K/SEC*
- f) Jitter, stability: *OK*
- g) Prime power: *+15 VOLTS*

60 db ATTEN
≈ 100 VOLTS/DIV
R_v ≈ 5.0 Ω
≈ 20 AMPS/DIV

+15 VOLTS
200 mA (max)