


PULSE GENERATOR  
PERFORMANCE CHECK

Model: *AVD-20-P*

S.N.: *5520*

Date: *JUNE 8 1990*



- a) Output signal amplitude: *0 TO +10 AMP*
- b) Pulse width: *TO 50*
- c) Rise time: *2 TO 50 NS*
- d) Fall time: *≤ 1 NS*
- e) PRF: *0 TO 20 KHz*
- f) Jitter, stability: *OK*

*50 46 ADDEN*

*2 32 00453 PIN + 15 VOLTS*

*R<sub>i</sub> ≈ 5.0 OH*

*≈ 6 AMP/PIN 400 mA (max)*

*PRF = 20 KHz*

*[Signature]*