


PULSE GENERATOR
PERFORMANCE CHECK

Model: AVR-A-1-PW-C-05-P

S.N.: 5521

Date: MAY 22 1990



- a) Output signal amplitude:
0 TO +200 V
- b) Pulse width:
50 NSEC TO 50 μ N
- c) Rise time:
1.0 NSEC
 ≤ 10 NSEC
- d) Fall time:
 ≤ 10 NSEC
- e) PRF:
0 TO 100 KHz
- f) Jitter, stability: (0.5% MAX DUTY CYCLE)
OK
- g) Prime power:

$R_L = 50 \Omega$
100 NSEC/DIV
50 VOLTS/DIV
PRF = 10 KHz

170/240 V.
50-60 KHz