


PULSE GENERATOR  
PERFORMANCE CHECK

Model: AVO-3A-P-1M

S.N.: 5524

Date: JUNE 8 1990



- a) Output signal amplitude: +25 TO +75 AMPS
- b) Pulse width: 2.0 ± 0.5 nSEC
- c) Rise time: NA
- d) Fall time: NA
- e) PRF: 0 TO 20 KHz
- f) Jitter, stability: OK
- g) Prime power: +15 VOLTS  
200 mA MAX

60 db ATTEN  
R<sub>L</sub> = 50 OHM  
100 VOLTS/DIV  
20 AMPS/DIV