


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVI-M17-N*
S.N.: *5557*
Date: *AUG 3 1990*



- a) Output signal amplitude:
71 40 VOLTS
- b) Pulse width:
2 TO 100 NSER
- c) Rise time:
≤ 350 PSEZ
- d) Fall time:
≤ 2 NSER
- e) PRF:
MIN DW: 0 TO 1000 Hz
MAX DW: 0 TO 20 KHz
- f) Jitter, stability:
OK
- g) Prime power:

50 dB ATTEN
32 VOLTS/DIV.

+15 ± 1.0 VOLT
100 mA (MAX)